



## VERIFICATION OF COMPLIANCE

*This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.*

**Technical Standard: EMC DIRECTIVE 93/42/EEC  
(EN 60601-1-2)**

*(Operation Environment: Medical Electrical Equipment)*

### General Information

**Applicant:** ARBOR Technology Corp.  
10F., No. 700, Zhong Zheng Rd., Zhong He,  
235 Taipei, Taiwan

### Product Description

**EUT Description:** Medical Tablet PC  
**Trade Name:** ARBOR Technology Corp.  
**Model Number:** M1255

### Measurement Standard

EN 60601-1-2: 2001 + A1: 2006  
EN 55011: 2007 (Group 1, Class B)  
EN 61000-3-2: 2006  
EN 61000-3-3: 1995 + A1: 2001 + A2: 2005  
(IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2006; IEC 61000-4-4: 2004;  
IEC 61000-4-5: 2005; IEC 61000-4-6: 2007; IEC 61000-4-8: 1993 + A1: 2000; IEC 61000-4-11: 2004)

### Measurement Facilities

**Sindian BU. Name:** Compliance Certification Services Inc.  
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan  
Tel: +886-2-22170894 / Fax: +886-2-22171029

*This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: 90427201-E*

Vince Chiang / Assistant Manager

Date: May 07, 2009



Compliance Certification Services Inc.






## VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the product designated below.

Product	Medical Tablet PC
Model	M1255
Trade name	ARBOR Technology Corp
Applicant	ARBOR Technology Corp. 10F., No. 700, Zhong Zheng Rd., Zhong He, 235 Taipei, Taiwan
Applicable Standard(s)	EN 60601-1(LVD): 1990 + A1: 1993 + A2: 1995 + A13: 1996 IEC 60601-1(LVD):1988 + A1:1991 + A2:1995
Report No.	90427201-LV
Test Laboratory	Compliance Certification Services Inc. No. 11, Wu-Gong 6th Rd., Wu-Gu Industrial Park Taipei Hsien, Taiwan Tel: +886-2-2299-9720 / Fax: +886-2-2299-1792

This device has been tested and found to comply with the stated standard(s), which is(are) required by the Council Directive of Medical Device Directive (93/42/EEC). The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

  
Kent Tseng / Safety Lab. Supervisor  
Date: June 17, 2009



Compliance Certification Services Inc.