CERTIFICATE OF COMPLIANCE

according to IEC / EN 60601-1-2:2001 +A1:2004 +A2:2005

EN 55011 : 1998 (Class B) IEC 61000-4-2 : 1995/ A1:1998/ A2:2000

EN 61000-3-2 : 2000/ A2: 2005 IEC 61000-4-3 : 2002/ A1:2002

EN 61000-3-3: 1995/ A1: 2001 IEC 61000-4-4: 2004

IEC 61000-4-5: 1995/ A1:2000 IEC 61000-4-6: 1996/ A1:2001 IEC 61000-4-8: 1993/ A1:2000

IEC 61000-4-11 : 2004

Applicant : ARBOR Technology Corp.

Address 10F., No. 700, Zhong Zheng Rd., Zhong He, 235

Taipei, Taiwan.

Equipment: 15" Medical Panel PC

Model No. : M1525

I HEREBY CERTIFY THAT:

The measurements shown in this test report were made in accordance with the procedures given in EUROPEAN COUNCIL DIRECTIVE 2004/108/EC. The equipment was *passed* the test performed according to European Standard EN 55011: 1998 (Class B), EN61000-3-2: 2000/ A2:2005, EN 61000-3-3:1995/ A1:2001 (IEC 61000-4-2: 1995/ A1:1998/ A2:2000, IEC 61000-4-3: 2002/ A1:2002, IEC 61000-4-4: 2004, IEC 61000-4-5: 1995/ A1:2000, IEC 61000-4-6: 1996/ A1:2001, IEC 61000-4-8: 1993/ A1:2000, IEC 61000-4-11: 2004).

The test was carried out on May. 26, 2008 at Exclusive Certification Corp.

Signature

Anson Chou / Manager

