



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: FCC Part 15 Class B (DoC)
IC ICES-003**

(Operation Environment: For Home And Office Use)

General Information

Applicant: ARBOR Technology Corp.
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan


Product Description

EUT Description: Panel PC
Trade Name: ARBOR
Model Number: M1726

Measurement Facilities

Sindian BU. Name: **Compliance Certification Services Inc.**
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T100330201-D



Vince Chiang / Assistant Manager

Date: April 30, 2010



Compliance Certification Services Inc.



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: EMC DIRECTIVE 2004/108/EC
(EN55022 / EN55024)**

(Operation Environment: Information Technology Equipment)

General Information

Applicant: ARBOR Technology Corp.
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan

Product Description

EUT Description: Panel PC
Trade Name: ARBOR
Model Number: M1726

Measurement Standard

EN 55022: 2006 + A1: 2007
EN 61000-3-2: 2006
EN 61000-3-3: 1995 + A1: 2001 + A2: 2005
EN 55024: 1998 + A1: 2001 + A2: 2003
(IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; IEC 61000-4-3: 2006; IEC 61000-4-4: 2004;
IEC 61000-4-5: 2005; IEC 61000-4-6: 2003 + A1: 2004 + A2: 2006; IEC 61000-4-8: 1993 + A1: 2000;
IEC 61000-4-11: 2004)

Measurement Facilities

Sindian BU. Name: Compliance Certification Services Inc.
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T100330201-E

Sam Wu for

Vince Chiang / Assistant Manager

Date: April 30, 2010



Compliance Certification Services Inc.



VERIFICATION OF COMPLIANCE

This Verification of Compliance is hereby issued to the below named company. The test results of this report relate only to the tested sample identified in this report.

**Technical Standard: EMC DIRECTIVE 93/42/EEC
(EN 60601-1-2)**

(Operation Environment: Medical Electrical Equipment)

General Information

Applicant: ARBOR Technology Corp.
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan

Product Description

EUT Description: Panel PC
Trade Name: ARBOR
Model Number: M1726

Measurement Standard

IEC 60601-1-2: 2007
EN 60601-1-2: 2007
CISPR 11: 2003 + A1: 2004
EN 55011: 2007 + A2: 2007 (Group 1, Class B)
IEC 61000-3-2: 2005; EN 61000-3-2: 2006,
IEC 61000-3-3: 1994 + A1: 2001 + A2: 2005; EN 61000-3-3: 1995 + A1: 2001 + A2: 2005
IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000; EN 61000-4-2: 1995 + A1: 1998 + A2: 2001
IEC 61000-4-3: 2006; EN 61000-4-3: 2006; IEC 61000-4-4: 2004; EN 61000-4-4: 2004
IEC 61000-4-5: 2005; EN 61000-4-5: 2006
IEC 61000-4-6: 2003 + A1: 2004 + A2: 2006; EN 61000-4-6: 2007
IEC 61000-4-8: 1993 + A1: 2000; EN 61000-4-8: 1993 + A1: 2001
IEC 61000-4-11: 2004; EN 61000-4-11: 2004

Measurement Facilities

Sindian BU. Name: Compliance Certification Services Inc.
No.163-1, Jhongsheng Rd., Sindian City, Taipei County 23151, Taiwan
Tel: +886-2-22170894 / Fax: +886-2-22171029

This device has been shown to be in compliance with and was tested in accordance with the measurement procedures specified in the Standards & Specifications listed above and as indicated in the measurement report number: T100330201-E

Sam Lin for

Vince Chiang / Assistant Manager

Date: April 30, 2010



Compliance Certification Services Inc.



1 TEST CERTIFICATION

Product: Panel PC

Model: M1726

Brand: ARBOR

Applicant: **ARBOR Technology Corp.**
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan

Manufacturer: **ARBOR Technology Corp.**
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan

Tested: April 1, 2010 ~ April 29, 2010

Applicable Standards: IEC 60601-1-2: 2007
EN 60601-1-2: 2007
CISPR 11: 2003 + A1: 2004
EN 55011: 2007 +A2: 2007 (Group 1, Class B)
IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000;
EN 61000-4-2: 1995 + A1: 1998 + A2: 2001
IEC 61000-4-3: 2006; EN 61000-4-3: 2006
IEC 61000-4-4: 2004; EN 61000-4-4: 2004
IEC 61000-4-5: 2005; EN 61000-4-5: 2006
IEC 61000-4-6: 2003 + A1: 2004 + A2: 2006; EN 61000-4-6: 2007
IEC 61000-4-8: 1993 + A1: 2000; EN 61000-4-8: 1993 + A1: 2001
IEC 61000-4-11: 2004; EN 61000-4-11: 2004
IEC 61000-3-2: 2005; EN 61000-3-2: 2006, Class D
IEC 61000-3-3: 1994 + A1: 2001 + A2: 2005;
EN 61000-3-3: 1995 + A1: 2001 + A2: 2005


Deviation from Applicable Standard

None

The above equipment has been tested by Compliance Certification Services Inc., and found compliance with the requirements set forth in the Directive 93/42/EEC concerning medical devices and technical standards mentioned above. The results of testing in this report apply only to the product/system, which was tested. Other similar equipment will not necessarily produce the same results due to production tolerance and measurement uncertainties.

Approved by:

Reviewed by:



Sam Hu
Section Manager



Vesta Hsu
Supervisor of report document dept.



1 TEST CERTIFICATION

Product: Panel PC**Model:** M1726**Brand:** ARBOR**Applicant:** **ARBOR Technology Corp.**
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan**Manufacturer:** **ARBOR Technology Corp.**
10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan**Tested:** April 1, 2010 ~ April 29, 2010**Applicable Standards:** **EN 55022: 2006 + A1: 2007, Class B****EN 61000-3-2: 2006****EN 61000-3-3: 1995 + A1: 2001 + A2: 2005****EN 55024: 1998 + A1: 2001 + A2: 2003**

IEC 61000-4-2: 1995 + A1: 1998 + A2: 2000

IEC 61000-4-3: 2006

IEC 61000-4-4: 2004

IEC 61000-4-5: 2005

IEC 61000-4-6: 2003 + A1: 2004 + A2: 2006

IEC 61000-4-8: 1993 + A1: 2000

IEC 61000-4-11: 2004

Deviation from Applicable Standard
None

The above equipment has been tested by Compliance Certification Services Inc., and found compliance with the requirements set forth in the technical standards mentioned above. The results of testing in this report apply only to the product/system, which was tested. Other similar equipment will not necessarily produce the same results due to production tolerance and measurement uncertainties.

Approved by:**Reviewed by:**

Sam Hu
Section Manager

Vesta Hsu
Supervisor of report document dept.



1 TEST RESULT CERTIFICATION

Product: Panel PC**Model:** M1726**Brand:** ARBOR**Applicant:** ARBOR Technology Corp.10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan**Manufacturer:** ARBOR Technology Corp.10F., No. 700, Zhong Zheng Rd., Zhong He, 235
Taipei, Taiwan**Tested:** April 1, 2010 ~ April 29, 2010

EMISSION			
Standard	Item	Result	Remarks
FCC 47 CFR Part 15 Subpart B, ICES-003 Issue 4 ANSI C63.4-2003	Conducted (Power Port)	PASS	Meet Class B limit
	Radiated	PASS	Meet Class B limit

- Note:** 1. The statements of test result on the above are decided by the request of test standard only; the measurement uncertainties are not factored into this compliance determination.
2. The information of measurement uncertainty is available upon the customer's request.

Deviation from Applicable Standard
None

The above equipment has been tested by Compliance Certification Services Inc., and found compliance with the requirements set forth in the technical standards mentioned above. The results of testing in this report apply only to the product/system, which was tested. Other similar equipment will not necessarily produce the same results due to production tolerance and measurement uncertainties.

Approved by:Sam Hu
Section Manager**Reviewed by:**Vesta Hsu
Supervisor of report document dept.